



DECLARATION OF CONFORMITY

This declaration is issued under the sole responsibility of the manufacturer INFICON. The object of the declaration is to certify that this equipment, designed and manufactured by:

INFICON Inc.
Two Technology Place
East Syracuse, NY 13057
USA

is in conformity with the relevant Community harmonization legislation. It has been constructed in accordance with good engineering practice in safety matters in force in the Community and does not endanger the safety of persons, domestic animals or property when properly installed and maintained and used in applications for which it was made.

Equipment Description: INFICON Transpector APXA, APXH, CPXH, SPXP Residual Gas Analyzer

Applicable Directives:	2014/30/EU (EMC)	Electromagnetic Compatibility Regulations
	2015/863/EU (RoHS)	Substances in Electrical and Electronic Equipment Regulations 2012 (AS AMENDED)

Applicable Standards:

Safety:	IEC 61010-1:2010/AMD1:2016 3 rd Edition: Listing # E115486
Emissions:	EN 61326-1:2020 (Radiated & Conducted Emissions)
	CISPR 11:2015/AMD2:2019
	EN 55011:2016/A11:2020 Emission Standard for Industrial, Scientific and Medical (ISM) radio RF equipment
	FCC Title 47 CFR Part 18 Class A emission requirements (USA)
RoHS:	EN IEC 63000: 2018
Immunity:	EN 61326:2020 (Industrial EMC Environments)
	(EMC – Measurement, Control & Laboratory Equipment) Immunity per Table 2
SEMI:	SEMI S2-0821 Compliant

CE Implementation Date: October 18, 2022

Authorized Representatives:

Paul Grimshaw
Paul Grimshaw
Compliance Manager, ISS

**Samuel
Carroll**

Digitally signed by
Samuel Carroll
Date: 2026.02.28
12:57:01 -05'00'

Samuel Carroll
Vice President of Engineering, ISS

ANY QUESTIONS RELATIVE TO THIS DECLARATION OR TO THE SAFETY OF INFICON'S PRODUCTS SHOULD BE DIRECTED, IN WRITING, TO THE AUTHORIZED REPRESENTATIVE AT THE ABOVE ADDRESS.