



**EU DECLARATION  
OF  
CONFORMITY**

This declaration is issued under the sole responsibility of the manufacturer INFICON. The object of the declaration is to certify that this equipment, designed and manufactured by:

**INFICON Inc.  
Two Technology Place  
East Syracuse, NY 13057  
USA**

is in conformity with the relevant Community harmonization legislation. It has been constructed in accordance with good engineering practice in safety matters in force in the Community and does not endanger the safety of persons, domestic animals or property when properly installed and maintained and used in applications for which it was made.

Equipment Description: IMM-100 Thin Film Deposition Monitor


Applicable Directives: 2014/35/EU (LVD)  
2014/30/EU (EMC)  
2011/65/EU (RoHS)

Applicable Standards:

|            |                                                                                                                                                                                                                                                                    |
|------------|--------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| Safety:    | EN 61010-1:2010 Safety Requirements for Electrical Equipment For Measurement, Control, and Laboratory Use.<br>Part 1: General Requirements                                                                                                                         |
| Emissions: | EN 61326-1:2013 (Radiated & Conducted Emissions)<br>(EMC – Measurement, Control & Laboratory Equipment)<br>EN 55011:2009/A1:2010 Group 1, Class A<br>ICES-001 Issue 4 ISM emissions requirements (Canada)<br>FCC 47 CFR Part 18 Class A emission requirement (USA) |
| Immunity:  | EN 61326-1:2013 (General EMC)<br>Class A: Immunity<br>(EMC – Measurement, Control & Laboratory Equipment)                                                                                                                                                          |
| RoHS:      | Fully compliant                                                                                                                                                                                                                                                    |

CE Implementation Date: August 23, 2019

Authorized Representatives:

  
Oliver Wyrsh  
President, ISS

  
Hjalmar Bruhns, Ph.D.  
VP of R&D, ISS

ANY QUESTIONS RELATIVE TO THIS DECLARATION OR TO THE SAFETY OF INFICON'S PRODUCTS SHOULD BE DIRECTED, IN WRITING, TO THE AUTHORIZED REPRESENTATIVE AT THE ABOVE ADDRESS.