

DECLARATION OF CONFORMITY

This is to certify that this equipment, designed and manufactured by INFICON meets the essential safety requirements of the European Union and is placed on the market accordingly. It has been constructed in accordance with good engineering practice in safety matters in force in the Community and does not endanger the safety of persons, domestic animals or property when properly installed and maintained and used in applications for which it was made.

Equipment Description: TEK-Mate® Refrigerant Leak Detector

Model Number: 705-202-Gxx/CNxx (Applicable to all Group numbers)

Applicable Directives: 2004/108/EC General EMC
2011/65/EU as amended by 2015/863/EU RoHS

Applicable Standards:

Safety:	EN 61010-1:2010	Safety requirements for electrical equipment for measurement, control, and laboratory use. General requirements.
Emissions:	EN 61326-1:2013	Edition 2.0 (Radiated, Conducted & Harmonic Emissions) (EMC – Measurement, Control & Laboratory Equipment)
	CISPR 11/EN 55011:2009 (+A1:2010)	Emission standard for industrial, scientific, and medical (ISM) radio RF equipment
Immunity:	EN 61326-1:2013	Edition 2.0 (EMC – Measurement, Control & Laboratory Equipment) Immunity per Table A.1 – Portable Test and Measurement
RoHS:	Compliant	
Other:	EN 14624:2012	Performance of portable leak detectors and of room monitors for halogenated refrigerants

CE Implementation Date: 18 August 2021



Manufacturer Representative
Brian King
INFICON
General Manager – Service Tools
Two Technology Place
East Syracuse, NY USA 13057
INFICON

EC Authorized Representative
INFICON GmbH
50968 Köln, Bonner Str. 498