

Declaration PSM 17.001
Date Issued November 2017
Requirements Measurement procedure above ground leak
detection, version 1.1

For above ground searching for gas leaks of gas grids, the Dutch guideline "Measurement procedure above ground leak detection, version 1.1". (Meetprocedure bovengronds lekzoeken versie 1.1) is used. Kiwa Technology has tested a specimen of the

IRwin gas detector

and declares that this specimen fulfils the requirements in this guideline.

For this purpose a prototype of the Irwin gas detector, with the following identification marks:

Type: MLD.130.030

S/N: 92000231

Year: 2016

is tested for the following aspects:

- Measurement technique
- Sensitivity
- Accuracy
- Stability
- Response time
- Purging time
- Cross sensitivity
- Dimensions and material of the carpet probe



ing. J.B.W. Wikkerink
Unit Manager Piping Systems & Materials
Kiwa Technology

Customer
INFICON GmbH
Bonner Str. 498
50968 Köln
Germany