

QUICK START GUIDE

How to set the Threshold Alarm at 100ppm on your IRwin Methane Leak Detector

1

Press  followed by 

2

Use  to navigate to *Thresholds* and press 



3

To select *CxHy* press 



4

If prompted, enter the Intermediate password code and repeat step 3. If not prompted, go ahead to the next step

5

Press  to open the drop-down menu. Use  to navigate to 100 and press 



6

Press  followed by  and 