

# QUICK START GUIDE


How to set the Start Mode to “Last Used” on the IRwin Methane Leak Detector

1


Press  followed by 

2

Use  to navigate to *General* and press 




3

Use  to navigate to *Misc* and press 



4




Use  to navigate to *Start Mode* and press 



5


If prompted, enter the Intermediate password code and repeat step 4. If not prompted, go ahead to the next step

6

Use  or  to navigate to *Last used* in the drop-down menu, press 



7

Press  followed by  and 