



防爆合格证

证号: GYJ16.1435X

由 INFICON AB
(地址: BOX 76, 581 02 Linköping, Sweden)

制造的产品:

名称 气体探测器

型号规格 IRwin SX_a

防爆标志 Ex ia II C T3 Ga

产品标准 /

图样编号 5574

经图样及技术文件的审查和样品检验, 确认上述产品符合 GB 3836.1-2010、GB 3836.4-2010、GB 3836.20-2010 标准, 特颁发此证。

本证书有效期: 2016年12月13日至2021年12月12日

- 备注
1. 安全使用注意事项见本证书附件。
 2. 证书编号后缀“X”表明产品具有安全使用特殊条件, 内容见本证书附件。
 3. [更改 1]增加新可选型号浇封剂(SI 5710) 2020年2月10日签发。

站长

国家级仪器仪表防爆安全监督检验站

颁发日期二〇一六年十二月十三日



本证书仅对与认可文件和样品一致的产品有效。

地址: 上海市漕宝路103号
邮编: 200233

网址: www.nepsi.org.cn
Email: info@nepsi.org.cn

电话: +86 21 64368180
传真: +86 21 64844580



EXPLOSION PROTECTION

CERTIFICATE OF CONFORMITY

Cert NO.GYJ16.1435X

This is to certify that the product

Gas Detector

manufactured by **INFICON AB**

(Address:BOX 76, 581 02 Linköping, Sweden)

which model is **IRwin SX_X**

Ex marking **Ex ia IIC T3 Ga**

product standard /

drawing number **5574**

has been inspected and certified by NEPSI, and that it conforms
to GB 3836.1-2010,GB 3836.4-2010,GB 3836.20-2010

This Approval shall remain in force until **2021.12.12**

Remarks 1.Conditions for safe use are specified in the attachment to this certificate.
2.Symbol "X" placed after the certification number denotes specific conditions of use,
which are specified in the attachment to this certificate.
3.[Variation I] Adding an alternative encapsulant SI 5710. Issued on 2020.02.10.

Director

National Supervision and Inspection Centre for
Explosion Protection and Safety of Instrumentation

Issued Date 2016.12.13



This Certificate is valid for products compatible with the documents and samples approved by NEPSI.

103 Cao Bao Road
Shanghai 200233, China

<http://www.nepsi.org.cn>
Email: info@nepsi.org.cn

Tel: +86 21 64368180
Fax: +86 21 64844580